



Electronic Temperature Instruments Ltd

Easting Close, Worthing, West Sussex BN14 8HQ · tel: 01903 202151 · fax: 01903 202445

specialists in the design and manufacture of temperature measuring equipment

Certificate of Conformity

E.T.I. Part Code: **825-505**

8051 Light Meter Lux/Foot Candle Meter

Specification:- Range: 0 to 30 000 lx; 0 to 2788 fc
Resolution: 1 Lux; 0.1 FC
Accuracy: $\pm 4\%$ or ± 50 digits
Made in China
Conforms to RoSH & REACH Legislation
Conforms to EC Reg 1935/2004 & EU Reg 10/2011

DECLARATION OF CONFORMITY

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

We also confirm that when used in an appropriate manner, the instruments are suitable for use within the Food Industry, and also, where appropriate and required, comply with EEC Directives with regard to European Legislation and CE Marking requirements. Including EC Regulation No 1935/2004 on Materials and articles intended to come into contact with food; & Commission Regulation (EU) No 10/2011 on plastic materials and articles intended to come into contact with food.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Electronic Temperature Instruments Ltd

David Carter
Commercial Manager Valid as of date of purchase.

